

CERTIFICATE OF ANALYSIS
Plasma/Serum Products
Coronavirus OC43 IgG Serum

Lot Number(s): See under "Lot/Donor Information"

Source Material: Serum

Preservative: -

Concentration of Marker/Test Method: See under "Test Results Data"

Expiry Date: None assigned. Users should determine suitability in their own system or refer to any known collection date

Fixed Data:

Product Code: H446-2

Antigen/Marker present: Coronavirus OC43 IgG

Species: Human

Format: Liquid

Storage Temperature: ≤-15°C. Minimise freeze-thaw cycles

Shipping Temperature: ≤-15°C recommended

Intended Use: For research and further manufacturing use

Health & Safety:

Infectious Disease Tests Performed: Material/donors have been tested for:

Material/donors tested for:	Result
Hepatitis B Surface Antigen	Negative/Non-reactive
Anti-HIV 1 & 2 Antibodies	Negative/Non-reactive
Anti-Hepatitis C Virus antibodies	Negative/Non-reactive

Chemicals: For further information please request a Safety Data Sheet.

Precaution: Handle as hazardous using appropriate Personal Protective Equipment. Refer to Safety Data Sheet (available on request) for safety information and precautions.

Test Results Data:


Test performed: Coronavirus OC43 IgG ELISA (OD/CO)
(Positive: >1.0)

Test result: See under "Lot/Donor information"

Lot/Donor information:

Lot#	Date of Collection	Donor Age	Donor Gender	Coronavirus OC43 IgG ELISA (OD/CO)
SE1485	19/05/2019	39	M	2.91
SE1486	08/07/2019	50	F	3.4
SE1487	10/04/2019	47	M	2.97
SE1488	03/07/2019	29	F	3.68
SE1489	21/04/2019	31	M	1.85

Quality Assurance Approval:

Signature:	
Date:	26 th Aug 2020